


Trade Name: Clip Remover  
(SANO-Clip Remover)




TYPE	Cat No.	Overall Length	Specific Clips
Straight 20.0cm	07-950-14	20.0cm	Sugita Aneurysm Clips of all types except LONG type.


**MATERIAL:**.....Jaws of Co-Cr alloy (ASTM F1058 G2,ISO 5832/7)  
Other parts stainless steel(AISI 410, 420)

**INSTRUCTION FOR USE:**

( Fig . 1 )



( Fig . 2 )



It is most important that the spring portion of clip is properly located within grooves of the remover jaws as illustrated.

If the spring portion of clip is not fully located within the jaws of the remover, it may slip or spring out.

**CAUTION:**.....SANO-Clip Remover is designed for use with Sugita clips only.  
Use only with the Sugita clips specified.  
Do not use with other manufactures clips.  
SANO-Clip Remover must be inspected before each use, if damaged do not use. Defective applying forceps can result in damage to the spring mechanism of a clip and alter the holding force.

\* **Name and address of manufacturer**



MIZUHO Corporation  
3-30-13 Hongo, Bunkyo-ku, Tokyo 113-0033, Japan  
<http://www.mizuho.co.jp>

\* **Authorized Representative Europe**



Emergo Europe B.V.  
Prinsessegracht 20, 2514 AP, The Hague  
The Netherlands